



Specification		102C1-50ST	Lot No	0104102C15026	Inspection date	2026.1.15
Size		50mL	Quantity	35000pcs	Report date	2026.1.15
Inspection standard		Q/321284CBB07			Sampling quantity	125pcs
Inspection Item		Technical performance requirements			Inspection result	
Biological properties	Sterility	Should be sterile			Conform	
Physical properties	Appearance	Transparent, smooth inner wall, enough strength			Conform	
	Size	Conform to standard			Conform	
	Leakage test	No leakage after water filling test			Conform	
	Clarity	It's clearly visible after the water level of the water container visual under normal light			Conform	
This lot of Centrifuge tube conform to Standard Q/321284CBB07						