

Certificate of Analysis

Cat No		102D0-2	Lot No	0104102D0226	Inspection date	2026.1.15
Size		1.8mL	Quantity	25000pcs	Report date	2026.1.15
Inspection standard		Q/321284CBB07			Sampling quantity	125pcs
Inspection Item		Technical performance requirements			Inspection result	
Physical properties	Appearance	Transparent, smooth inner wall, enough strength			Conform	
	Size	Conform to standard			Conform	
	Leakage test	No leakage after water filling test			Conform	
	Clarity	It's clearly visible after the water level of the water container visual under normal light			Conform	

This lot of Cryovial tube conform to Standard Q/321284CBB07