

Certificate of Analysis

Name		120mL Urine Collection Container	LOT NO	260130	Inspection Date	2026.2.2
Cat No		111A1-100	Sterile No	260130	Report Date	2026.2.2
Inspection Basis		Q/321284CB B07	Lot Qty	5000pcs	Sampling quantity	80pcs
Inspection Items		Technical Requirements			Inspection results	Judgement
Biological performance	Sterile	Should be sterile			Conform	Pass
Physical Performance	Appearance	The surface of the product should be smooth, pollution-free, and free from defects such as oil stains, burrs, scratches, missing materials, cracks, weld marks, etc.; The product scale should be clear; The product label should be intact.			Conform	Pass
	External dimensions	Should meet the requirements of the drawings.			Conform	Pass
	Tightness	Collection cup lid and cup body fitting part should not leak liquid.			Conform	Pass
	Nominal capacity	Add water to the nominal capacity mark, and the volume of water added should be between 90% and 110% of the nominal capacity.			Conform	Pass
Conclusion		After inspection, the batch of urine collection container meets the requirements of Q/321284CBB07 standard.				